

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
 Data File : VU045497.D
 Acq On : 01 Nov 2021 19:17
 Operator : SY/MD
 Sample : M4431-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-383-385

Quant Time: Nov 02 07:36:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	191309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	373842	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	371850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	181084	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	162239	52.090	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.180%	
35) Dibromofluoromethane	5.292	113	120711	50.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	7.903	98	480325	51.529	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.060%	
62) 4-Bromofluorobenzene	10.636	95	185516	52.229	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.460%	
Target Compounds						
16) Acetone	2.639	43	8099	5.728	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
Data File : VU045497.D
Acq On : 01 Nov 2021 19:17
Operator : SY/MD
Sample : M4431-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-383-385

Quant Time: Nov 02 07:36:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

