

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
 Data File : VU045502.D
 Acq On : 01 Nov 2021 21:12
 Operator : SY/MD
 Sample : M4431-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-458-460

Quant Time: Nov 02 07:37:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	186787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	371614	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	362230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	169201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	158851	52.237	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.480%
35) Dibromofluoromethane	5.295	113	119631	50.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	7.902	98	474189	51.176	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.360%
62) 4-Bromofluorobenzene	10.635	95	176307	49.933	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.860%
Target Compounds						
16) Acetone	2.642	43	3631	2.630	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
Data File : VU045502.D
Acq On : 01 Nov 2021 21:12
Operator : SY/MD
Sample : M4431-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-458-460

Quant Time: Nov 02 07:37:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

