

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
 Data File : VU045514.D
 Acq On : 02 Nov 2021 15:12
 Operator : SY/MD
 Sample : M4431-04RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-383-385RE

Quant Time: Nov 02 16:18:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	185504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	330846	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	365198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	170830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	148443	49.152	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 98.300%			
35) Dibromofluoromethane	5.292	113	116685	55.467	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.940%			
50) Toluene-d8	7.899	98	485659	58.873	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 117.740%#			
62) 4-Bromofluorobenzene	10.636	95	179918	57.235	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 114.480%			
Target Compounds						
16) Acetone	2.646	43	4377	3.192	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110221\
Data File : VU045514.D
Acq On : 02 Nov 2021 15:12
Operator : SY/MD
Sample : M4431-04RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-383-385RE

Quant Time: Nov 02 16:18:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

