

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110223\
 Data File : VU056056.D
 Acq On : 02 Nov 2023 09:57
 Operator : MD/SY
 Sample : VU1102WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK121

Quant Time: Nov 03 00:37:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 03 00:21:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	318216	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	306966	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	127912	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	110337	4.383	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.600%
7) Chloroethane-d5	1.914	69	89202	4.546	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.000%
11) 1,1-Dichloroethene-d2	2.567	65	48342	4.524	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.400%
20) 2-Butanone-d5	4.637	46	209623	46.679	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.360%
24) Chloroform-d	5.058	84	211240	4.529	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	5.698	65	99573	4.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.724	84	425485	4.653	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.689	67	130040	4.797	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.894	98	363550	4.383	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	8.177	79	45562	4.298	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.634	63	155250	57.688	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.380%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	90989	4.609	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	112961	4.733	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110223\
Data File : VU056056.D
Acq On : 02 Nov 2023 09:57
Operator : MD/SY
Sample : VU1102WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK121

Quant Time: Nov 03 00:37:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 03 00:21:27 2023
Response via : Initial Calibration

