

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U110320\
 Data File : VU041066.D
 Acq On : 03 Nov 2020 10:11
 Operator : SY/MD
 Sample : VU1103WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK14

Quant Time: Nov 04 01:05:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 04 00:51:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	50139	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	51634	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	21497	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	21707	5.15	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.00%
7) Chloroethane-d5	1.92	69	17239	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	2.58	63	32110	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	4.66	46	86973	62.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.40%
24) Chloroform-d	5.08	84	38120	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.71	65	25507	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.40%
32) Benzene-d6	5.74	84	73115	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.70	67	24561	5.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.20%
41) Toluene-d8	7.90	98	59024	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	8.18	79	10008	5.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.60%
46) 2-Hexanone-d5	8.64	63	60743	57.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.78%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	22872	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	22217	5.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110320\
Data File : VU041066.D
Acq On : 03 Nov 2020 10:11
Operator : SY/MD
Sample : VU1103WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK14

Quant Time: Nov 04 01:05:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 04 00:51:13 2020
Response via : Initial Calibration

