

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110322\
 Data File : VU051742.D
 Acq On : 03 Nov 2022 18:40
 Operator : JC/MD
 Sample : N5407-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW4Y3

Quant Time: Nov 04 00:53:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 04 00:49:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.347	114	162250	5.000	ug/L	0.10
28) Chlorobenzene-d5	9.440	117	168781	5.000	ug/L	0.02
58) 1,4-Dichlorobenzene-d4	11.816	152	80976	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.662	65	53283	4.360	ug/L	0.06
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.200%
7) Chloroethane-d5	1.987	69	52399	4.112	ug/L	0.07
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.200%
11) 1,1-Dichloroethene-d2	2.655	65	25279	5.700	ug/L	0.09
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.000%
20) 2-Butanone-d5	4.771	46	201961	45.635	ug/L	0.15
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.280%
24) Chloroform-d	5.144	84	123092	4.305	ug/L	0.08
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.816	65	60932	3.945	ug/L	0.12
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
32) Benzene-d6	5.835	84	224467	4.214	ug/L	0.11
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.774	67	77512	4.162	ug/L	0.08
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.200%
41) Toluene-d8	7.945	98	167646	4.003	ug/L	0.05
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.000%
43) trans-1,3-Dichloroprop...	8.221	79	23425	4.113	ug/L	0.04
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.671	63	116056	33.854	ug/L	0.04
Spiked Amount	50.000	Range	45 - 130	Recovery	=	67.700%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	68642	4.022	ug/L	0.01
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	70965	4.775	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.887	76	31133	0.685	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110322\
Data File : VU051742.D
Acq On : 03 Nov 2022 18:40
Operator : JC/MD
Sample : N5407-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW4Y3

Quant Time: Nov 04 00:53:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 04 00:49:19 2022
Response via : Initial Calibration

