



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110322\
Data File : VU051723.D
Acq On : 03 Nov 2022 09:31
Operator : JC/MD
Sample : BFB163
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB163

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Nov 04 00:49:19 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	39243	PASS
75	95	30	80	48.3	113869	PASS
95	95	100	100	100.0	235627	PASS
96	95	5	9	6.6	15565	PASS
173	174	0.00	2	1.2	2380	PASS
174	95	50	120	81.4	191787	PASS
175	174	5	9	7.0	13485	PASS
176	174	95	101	95.4	182912	PASS
177	176	5	9	6.5	11941	PASS

m/z	Abundance
39.90	714.0
43.90	435.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

m/z	Abundance
36.00	1664.0
37.00	8225.0
38.00	8698.0
39.00	4308.0
40.00	1131.0
44.00	1651.0
45.00	1734.0
46.00	344.0
47.00	2542.0
48.00	1559.0
49.00	8306.0
50.00	39032.0
51.00	13215.0
52.00	645.0
52.90	246.0
54.00	104.0
55.00	544.0
56.00	3102.0
57.00	6097.0
58.00	258.0
60.00	2296.0
61.00	11295.0
62.00	10974.0
63.10	9494.0
64.00	2024.0
65.00	2958.0
66.00	270.0
66.90	615.0
68.00	21936.0
69.00	21992.0
70.00	1682.0
72.00	1162.0
73.00	9913.0
74.00	37944.0
75.00	111400.0
76.00	10154.0
77.00	1330.0
78.00	1028.0
78.90	8106.0
80.00	2436.0
80.90	8201.0
81.90	2160.0
83.00	159.0
85.80	358.0
87.00	10543.0
88.00	9763.0
90.90	846.0
92.00	5997.0
93.00	10312.0
94.00	26104.0
95.00	229440.0
96.00	15434.0
97.00	514.0
104.00	981.0
104.90	306.0
105.90	878.0
106.80	197.0
107.80	126.0
110.90	172.0
111.80	152.0
112.90	184.0
114.90	191.0
115.90	779.0
116.90	1819.0
118.00	803.0
118.90	1506.0
127.90	972.0
128.80	289.0
129.90	843.0
130.80	291.0
135.00	396.0
136.90	388.0
139.90	165.0
140.90	1926.0
141.90	176.0
143.00	3688.0
143.90	197.0
145.00	1422.0
145.90	316.0
146.90	216.0
147.90	562.0
148.80	132.0
150.00	150.0
151.80	112.0
152.90	180.0
154.00	127.0
155.00	450.0
156.90	417.0
158.90	285.0
160.80	209.0
171.00	582.0
172.00	3188.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

173.00	2139.0
174.00	182400.0
175.00	13352.0
176.00	173056.0
177.00	11225.0
177.90	315.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

m/z	Abundance
36.00	1586.0
37.00	8774.0
38.00	8760.0
39.00	4342.0
40.00	1110.0
44.00	1619.0
45.00	1830.0
46.10	315.0
47.00	2607.0
48.00	1685.0
49.00	8677.0
50.00	40440.0
51.00	13711.0
52.00	688.0
52.90	227.0
54.00	127.0
55.00	571.0
56.00	3158.0
57.10	6236.0
58.00	257.0
60.00	2415.0
61.00	11509.0
62.00	11261.0
63.10	9948.0
64.00	1911.0
65.00	3129.0
66.00	344.0
67.00	584.0
68.00	22808.0
69.00	23416.0
70.00	1811.0
72.00	1163.0
73.00	10239.0
74.00	40008.0
75.00	117752.0
76.00	10664.0
77.00	1423.0
78.00	1021.0
78.90	8185.0
80.00	2646.0
81.00	8368.0
81.90	2162.0
83.00	190.0
85.90	411.0
87.00	11249.0
88.00	10221.0
90.90	845.0
92.00	6748.0
93.00	10964.0
94.00	27928.0
95.00	242624.0
96.00	16054.0
97.00	473.0
104.00	917.0
104.90	270.0
105.90	837.0
106.90	216.0
108.00	154.0
110.90	197.0
111.90	137.0
112.90	180.0
114.90	232.0
115.90	870.0
116.90	1989.0
117.90	1020.0
118.90	1564.0
127.90	947.0
128.80	273.0
129.90	924.0
130.90	304.0
135.00	421.0
137.00	381.0
139.80	152.0
140.90	2166.0
141.90	211.0
143.00	3767.0
143.90	260.0
145.00	1460.0
145.90	386.0
147.00	216.0
148.00	554.0
149.00	120.0
150.00	204.0
151.90	119.0
152.90	233.0
154.00	143.0
155.00	559.0
156.90	436.0
158.90	240.0
160.90	258.0
171.00	638.0
172.00	3269.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

173.00	2452.0
174.00	197504.0
175.00	13907.0
176.00	188672.0
177.00	12310.0
178.00	321.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

m/z	Abundance
36.00	1438.0
37.00	8597.0
38.00	8044.0
39.00	3911.0
40.00	979.0
44.00	1439.0
45.00	1655.0
46.10	292.0
47.00	2350.0
48.00	1576.0
49.00	8053.0
50.00	38256.0
51.00	12683.0
52.00	598.0
53.10	210.0
54.10	120.0
55.00	502.0
56.00	2996.0
57.10	5696.0
58.00	196.0
60.00	2260.0
61.00	10387.0
62.00	10810.0
63.10	9393.0
64.00	1654.0
65.00	2868.0
66.00	300.0
67.00	504.0
68.00	21824.0
69.00	22168.0
70.00	1698.0
72.00	1068.0
73.00	9506.0
74.00	38160.0
75.00	112456.0
76.00	9796.0
77.00	1357.0
78.00	896.0
79.00	7539.0
80.00	2644.0
81.00	7826.0
81.90	1985.0
83.00	173.0
85.90	442.0
87.00	11031.0
88.00	9925.0
90.90	737.0
92.00	6578.0
93.00	10516.0
94.00	27088.0
95.00	234816.0
96.10	15206.0
97.00	422.0
102.90	102.0
103.90	798.0
104.90	266.0
105.90	780.0
106.80	225.0
108.00	217.0
111.00	215.0
111.90	115.0
112.90	167.0
114.80	256.0
115.90	869.0
117.00	1923.0
117.90	1038.0
118.90	1474.0
128.00	757.0
128.80	248.0
129.90	805.0
131.00	301.0
135.00	348.0
137.00	336.0
139.80	111.0
140.90	2154.0
142.00	199.0
143.00	3431.0
143.90	326.0
145.00	1359.0
145.90	419.0
146.90	176.0
148.00	539.0
149.00	104.0
150.00	201.0
152.90	229.0
154.00	146.0
155.00	608.0
157.00	401.0
158.90	201.0
160.90	285.0
170.90	625.0
172.00	3048.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163

173.00	2550.0
174.00	195456.0
175.00	13195.0
176.00	187008.0
177.00	12288.0
178.00	281.0

Instrument :
MSVOA_U
ClientSampleId :
BFB163