

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110323\
 Data File : VU056075.D
 Acq On : 03 Nov 2023 12:25
 Operator : MD/SY
 Sample : VU1103WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK122

Quant Time: Nov 04 01:40:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 04 01:40:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	347137	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	337246	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	149004	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	98402	3.583	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.600%	
7) Chloroethane-d5	1.914	69	84315	3.939	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	78.800%	
11) 1,1-Dichloroethene-d2	2.567	65	43007	3.689	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.800%	
20) 2-Butanone-d5	4.634	46	224476	45.822	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.640%	
24) Chloroform-d	5.062	84	208292	4.093	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.800%	
26) 1,2-Dichloroethane-d4	5.701	65	99468	4.012	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.200%	
32) Benzene-d6	5.724	84	418068	4.162	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
36) 1,2-Dichloropropane-d6	6.685	67	125442	4.212	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	84.200%	
41) Toluene-d8	7.894	98	350902	3.851	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.000%	
43) trans-1,3-Dichloroprop...	8.177	79	46187	3.965	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.400%	
46) 2-Hexanone-d5	8.631	63	165476	55.967	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	111.940%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	97423	4.492	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	126051	4.534	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.600%	
Target Compounds						
16) Methylene chloride	3.049	84	10289	0.540	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110323\
Data File : VU056075.D
Acq On : 03 Nov 2023 12:25
Operator : MD/SY
Sample : VU1103WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK122

Quant Time: Nov 04 01:40:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 04 01:40:01 2023
Response via : Initial Calibration

