

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045536.D
 Acq On : 03 Nov 2021 21:42
 Operator : SY/MD
 Sample : PB140448ZHE#02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140448ZHE#02

Quant Time: Nov 05 13:53:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	178410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	279732	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	347273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	166606	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	145205	49.992	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.980%
35) Dibromofluoromethane	5.292	113	105655	59.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.800%
50) Toluene-d8	7.899	98	472084	67.684	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	135.360%#
62) 4-Bromofluorobenzene	10.636	95	155619	58.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.100%
Target Compounds						
					Qvalue	
16) Acetone	2.646	43	28716	21.777	ug/l	97
20) Methylene Chloride	3.051	84	9421	3.585	ug/l	96
43) Isopropyl Acetate	5.909	43	24552	5.052	ug/l	98
70) Styrene	10.121	104	444	1.029	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045536.D
Acq On : 03 Nov 2021 21:42
Operator : SY/MD
Sample : PB140448ZHE#02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140448ZHE#02

Quant Time: Nov 05 13:53:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

