

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045537.D
 Acq On : 03 Nov 2021 22:05
 Operator : SY/MD
 Sample : PB140448ZHE#03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140448ZHE#03

Quant Time: Nov 05 13:54:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	174447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	352952	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	270545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	139593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	123264	43.402	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.800%		
35) Dibromofluoromethane	5.295	113	105232	46.890	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.780%		
50) Toluene-d8	7.900	98	323643	36.775	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	73.560%#		
62) 4-Bromofluorobenzene	10.636	95	124321	37.072	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	74.140%#		
Target Compounds						
16) Acetone	2.646	43	27914	21.650	ug/l	96
20) Methylene Chloride	3.054	84	9705	3.777	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045537.D
Acq On : 03 Nov 2021 22:05
Operator : SY/MD
Sample : PB140448ZHE#03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140448ZHE#03

Quant Time: Nov 05 13:54:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

