

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045538.D
 Acq On : 03 Nov 2021 22:28
 Operator : SY/MD
 Sample : PB140448ZHE#04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140448ZHE#04

Quant Time: Nov 05 13:54:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	182759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	362477	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	368771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	170457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	144705	48.634	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.295	113	110772	48.061	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	7.902	98	483982	53.550	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.100%
62) 4-Bromofluorobenzene	10.635	95	182429	52.970	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.940%
Target Compounds						
					Qvalue	
16) Acetone	2.649	43	24670	18.264	ug/l	95
20) Methylene Chloride	3.054	84	9952	3.697	ug/l	97
43) Isopropyl Acetate	5.915	43	24846	3.945	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045538.D
Acq On : 03 Nov 2021 22:28
Operator : SY/MD
Sample : PB140448ZHE#04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140448ZHE#04

Quant Time: Nov 05 13:54:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

