

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045540.D
 Acq On : 03 Nov 2021 23:14
 Operator : SY/MD
 Sample : PB140448ZHE#06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB140448ZHE#06

Quant Time: Nov 05 13:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	181489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	366215	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	347825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	176483	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	144402	48.872	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.740%
35) Dibromofluoromethane	5.292	113	113842	48.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.780%
50) Toluene-d8	7.899	98	321205	35.177	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	70.360%#
62) 4-Bromofluorobenzene	10.636	95	176125	50.617	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.240%
Target Compounds						
					Qvalue	
16) Acetone	2.649	43	12553	9.358	ug/l	95
20) Methylene Chloride	3.054	84	6285	2.351	ug/l #	90
43) Isopropyl Acetate	5.909	43	25629	4.028	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045540.D
Acq On : 03 Nov 2021 23:14
Operator : SY/MD
Sample : PB140448ZHE#06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB140448ZHE#06

Quant Time: Nov 05 13:54:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

