

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045541.D
 Acq On : 03 Nov 2021 23:36
 Operator : SY/MD
 Sample : M4470-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SP-3

Quant Time: Nov 05 13:57:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	145137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	264567	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	264694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	135000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	86784	36.728	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	73.460%#
35) Dibromofluoromethane	5.295	113	78473	46.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.300%
50) Toluene-d8	7.899	98	320701	48.615	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	10.636	95	121131	48.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.380%
Target Compounds						
					Qvalue	
16) Acetone	2.649	43	12526	11.677	ug/l	90
20) Methylene Chloride	3.051	84	5955	2.786	ug/l #	84
43) Isopropyl Acetate	5.912	43	13110	2.852	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045541.D
Acq On : 03 Nov 2021 23:36
Operator : SY/MD
Sample : M4470-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SP-3

Quant Time: Nov 05 13:57:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

