

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045544.D
 Acq On : 04 Nov 2021 00:45
 Operator : SY/MD
 Sample : M4472-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-JS

Quant Time: Nov 05 13:58:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	155461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	288309	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	285074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	149080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	96347	38.068	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	76.140%#
35) Dibromofluoromethane	5.295	113	86093	46.963	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	7.899	98	351759	48.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	10.636	95	132980	48.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.080%
Target Compounds						
					Qvalue	
16) Acetone	2.646	43	25522	22.212	ug/l	91
20) Methylene Chloride	3.054	84	8087	3.532	ug/l #	85
43) Isopropyl Acetate	5.912	43	15301	3.055	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045544.D
Acq On : 04 Nov 2021 00:45
Operator : SY/MD
Sample : M4472-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-JS

Quant Time: Nov 05 13:58:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

