

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
 Data File : VU045546.D
 Acq On : 04 Nov 2021 01:31
 Operator : SY/MD
 Sample : M4472-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-KD

Quant Time: Nov 05 13:58:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 29 11:09:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	153134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	291496	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	340642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	174695	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	95799	38.426	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	76.860%#
35) Dibromofluoromethane	5.292	113	85606	46.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	7.900	98	361797	49.778	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	10.636	95	180607	65.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	130.420%#
Target Compounds						
					Qvalue	
16) Acetone	2.652	43	14013	12.381	ug/l	95
20) Methylene Chloride	3.051	84	7129	3.161	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110421\
Data File : VU045546.D
Acq On : 04 Nov 2021 01:31
Operator : SY/MD
Sample : M4472-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-KD

Quant Time: Nov 05 13:58:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U102821W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 29 11:09:42 2021
Response via : Initial Calibration

