

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110518\
 Data File : VU027948.D
 Acq On : 05 Nov 2018 11:40
 Operator : MD/SY
 Sample : PB114473ZHE#03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114473ZHE#03

Quant Time: Nov 06 04:05:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	205057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	371714	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	334397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	145077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	161020	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
35) Dibromofluoromethane	4.89	113	123538	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
50) Toluene-d8	7.57	98	472230	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	10.31	95	163599	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
Target Compounds						
16) Acetone	2.32	43	10298	5.444	ug/l	97
20) Methylene Chloride	2.71	84	6891	2.495	ug/l	99
43) Isopropyl Acetate	5.55	43	241861	27.063	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110518\
Data File : VU027948.D
Acq On : 05 Nov 2018 11:40
Operator : MD/SY
Sample : PB114473ZHE#03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114473ZHE#03

Quant Time: Nov 06 04:05:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

