

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110518\
 Data File : VU027950.D
 Acq On : 05 Nov 2018 12:26
 Operator : MD/SY
 Sample : PB114473ZHE#05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114473ZHE#05

Quant Time: Nov 06 04:08:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	360507	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	325298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	141467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	158280	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
35) Dibromofluoromethane	4.89	113	116899	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
50) Toluene-d8	7.57	98	461399	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	10.31	95	160301	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
Target Compounds						
16) Acetone	2.32	43	10612	5.825	ug/l	99
18) Methyl Acetate	2.62	43	5118	1.021	ug/l #	87
20) Methylene Chloride	2.70	84	6987	2.627	ug/l	98
43) Isopropyl Acetate	5.55	43	233960	26.993	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110518\
Data File : VU027950.D
Acq On : 05 Nov 2018 12:26
Operator : MD/SY
Sample : PB114473ZHE#05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114473ZHE#05

Quant Time: Nov 06 04:08:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

