

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110518\
 Data File : VU027963.D
 Acq On : 05 Nov 2018 17:28
 Operator : MD/SY
 Sample : PB114473ZHE#18
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114473ZHE#18

Quant Time: Nov 06 04:40:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	193461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	356231	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	328096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	141635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	158367	55.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.12%	
35) Dibromofluoromethane	4.89	113	118742	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
50) Toluene-d8	7.57	98	464154	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
62) 4-Bromofluorobenzene	10.31	95	161449	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
Target Compounds						
16) Acetone	2.32	43	9292	5.207	ug/l	97
18) Methyl Acetate	2.62	43	5336	1.087	ug/l	94
20) Methylene Chloride	2.71	84	6034	2.316	ug/l #	83
43) Isopropyl Acetate	5.55	43	224538	26.217	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110518\
Data File : VU027963.D
Acq On : 05 Nov 2018 17:28
Operator : MD/SY
Sample : PB114473ZHE#18
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114473ZHE#18

Quant Time: Nov 06 04:40:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

