

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110518\
 Data File : VU027964.D
 Acq On : 05 Nov 2018 17:51
 Operator : MD/SY
 Sample : PB114473ZHE#19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114473ZHE#19

Quant Time: Nov 06 04:42:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	354673	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	317960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	139214	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	156283	55.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.80%	
35) Dibromofluoromethane	4.89	113	116099	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
50) Toluene-d8	7.57	98	456080	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
62) 4-Bromofluorobenzene	10.31	95	155953	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
Target Compounds						
16) Acetone	2.32	43	9211	5.263	ug/l	98
18) Methyl Acetate	2.62	43	5143	1.068	ug/l	97
20) Methylene Chloride	2.70	84	6158	2.410	ug/l	90
43) Isopropyl Acetate	5.55	43	229757	26.944	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110518\
Data File : VU027964.D
Acq On : 05 Nov 2018 17:51
Operator : MD/SY
Sample : PB114473ZHE#19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114473ZHE#19

Quant Time: Nov 06 04:42:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

