

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110518\
 Data File : VU027966.D
 Acq On : 05 Nov 2018 18:37
 Operator : MD/SY
 Sample : J5816-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-4A-D

Quant Time: Nov 06 04:45:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	189499	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	346625	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	307612	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	128134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	154287	54.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.52%	
35) Dibromofluoromethane	4.89	113	114438	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
50) Toluene-d8	7.57	98	445468	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	10.31	95	146988	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
Target Compounds						
8) Diethyl Ether	2.11	74	13028	9.110	ug/l	94
16) Acetone	2.32	43	40872	23.382	ug/l	98
18) Methyl Acetate	2.62	43	5563	1.157	ug/l #	88
20) Methylene Chloride	2.70	84	40301	15.791	ug/l	97
43) Isopropyl Acetate	5.55	43	195298	23.435	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110518\
Data File : VU027966.D
Acq On : 05 Nov 2018 18:37
Operator : MD/SY
Sample : J5816-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-4A-D

Quant Time: Nov 06 04:45:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

