

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU110519\
 Data File : VU035629.D
 Acq On : 05 Nov 2019 14:38
 Operator : JC/SP
 Sample : K5679-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CM4

Quant Time: Nov 06 05:51:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 06 03:03:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	643771	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	618339	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	186069	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	183936	35.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.76%
7) Chloroethane-d5	1.94	69	168726	40.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.26%
11) 1,1-Dichloroethene-d2	2.60	63	244251	30.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.42%
21) 2-Butanone-d5	4.69	46	286057	94.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.18%
24) Chloroform-d	5.11	84	365358	45.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.75	65	229744	46.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.40%
32) Benzene-d6	5.77	84	756928	46.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.92%
36) 1,2-Dichloropropane-d6	6.74	67	250633	47.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.90%
41) Toluene-d8	7.93	98	681731	44.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.68%
43) trans-1,3-Dichloropropene-	8.21	79	109561	45.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.00%
47) 2-Hexanone-d5	8.68	63	184980	92.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.34%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	344647	46.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.74%
64) 1,2-Dichlorobenzene-d4	12.22	152	188613	50.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.52%


Target Compounds					Qvalue
13) Acetone	2.67	43	11874	4.444 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU110519\
Data File : VU035629.D
Acq On : 05 Nov 2019 14:38
Operator : JC/SP
Sample : K5679-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CM4

Quant Time: Nov 06 05:51:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 06 03:03:41 2019
Response via : Initial Calibration





#13

Acetone

Concen: 4.444 ug/L

RT: 2.67 min Scan# 413

Delta R.T. 0.00 min

Lab File: VU035629.D

Acq: 05 Nov 2019 14:38

Instrument :

MSVOA_U

ClientSampleId :

C0CM4

Tgt Ion: 43 Resp: 11874

Ion Ratio Lower Upper

43 100

58 35.3 0.0 70.0

