

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U110520\
 Data File : VU041113.D
 Acq On : 05 Nov 2020 18:35
 Operator : SY/MD
 Sample : L4648-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB4

Quant Time: Nov 06 02:30:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 06 01:44:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	50922	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	52916	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	22785	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	16542	3.87	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.40%
7) Chloroethane-d5	1.92	69	13824	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.20%
11) 1,1-Dichloroethene-d2	2.58	63	23212	2.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.40%#
20) 2-Butanone-d5	4.65	46	83132	59.01	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.02%
24) Chloroform-d	5.08	84	37860	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
26) 1,2-Dichloroethane-d4	5.72	65	24113	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.80%
32) Benzene-d6	5.75	84	65232	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
36) 1,2-Dichloropropane-d6	6.70	67	23528	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.91	98	50093	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.20%
43) trans-1,3-Dichloropropene-	8.19	79	9118	4.56	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.20%
46) 2-Hexanone-d5	8.64	63	57432	53.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.82%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	22944	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	21447	5.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.20%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	826	0.170 ug/L	

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110520\
Data File : VU041113.D
Acq On : 05 Nov 2020 18:35
Operator : SY/MD
Sample : L4648-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB4

Quant Time: Nov 06 02:30:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 06 01:44:25 2020
Response via : Initial Calibration

