

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045561.D
 Acq On : 05 Nov 2021 16:12
 Operator : SY/MD
 Sample : M4475-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP02

Quant Time: Nov 05 17:05:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	136128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	260361	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	262699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	129609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	114399	56.114	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.220%
35) Dibromofluoromethane	5.292	113	88061	50.659	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	7.899	98	346020	53.532	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.060%
62) 4-Bromofluorobenzene	10.636	95	130817	50.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
16) Acetone	2.626	43	6868	6.618	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045561.D
Acq On : 05 Nov 2021 16:12
Operator : SY/MD
Sample : M4475-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP02

Quant Time: Nov 05 17:05:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

