

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045583.D
 Acq On : 06 Nov 2021 01:22
 Operator : SY/MD
 Sample : M4508-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-538-540

Quant Time: Nov 06 21:50:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	126859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	242831	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	244343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	122510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	107881	56.784	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.560%	
35) Dibromofluoromethane	5.292	113	82919	51.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.280%	
50) Toluene-d8	7.899	98	320105	53.098	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.200%	
62) 4-Bromofluorobenzene	10.635	95	121797	50.382	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.760%	
Target Compounds						
16) Acetone	2.629	43	3031	3.134	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045583.D
Acq On : 06 Nov 2021 01:22
Operator : SY/MD
Sample : M4508-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-538-540

Quant Time: Nov 06 21:50:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

