

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045584.D
 Acq On : 06 Nov 2021 01:45
 Operator : SY/MD
 Sample : M4508-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-558-560

Quant Time: Nov 06 21:51:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	124361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	240043	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	242292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	122101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	107167	57.541	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 115.080%	
35) Dibromofluoromethane	5.295	113	82456	51.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.900%	
50) Toluene-d8	7.899	98	316440	53.099	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.200%	
62) 4-Bromofluorobenzene	10.635	95	120122	50.267	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.540%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045584.D
Acq On : 06 Nov 2021 01:45
Operator : SY/MD
Sample : M4508-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-558-560

Quant Time: Nov 06 21:51:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

