

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045585.D
 Acq On : 06 Nov 2021 02:08
 Operator : SY/MD
 Sample : M4508-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-VPB-RW9-GW-583-585

Quant Time: Nov 06 21:51:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	125628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	239740	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	243350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	121624	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	106952	56.846	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.700%
35) Dibromofluoromethane	5.295	113	80749	50.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.900%
50) Toluene-d8	7.899	98	316500	53.176	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.360%
62) 4-Bromofluorobenzene	10.635	95	119490	50.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045585.D
Acq On : 06 Nov 2021 02:08
Operator : SY/MD
Sample : M4508-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-VPB-RW9-GW-583-585

Quant Time: Nov 06 21:51:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

