

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045588.D
 Acq On : 06 Nov 2021 03:17
 Operator : SY/MD
 Sample : M4478-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TWP02

Quant Time: Nov 06 21:52:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	122437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	237414	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	238896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	118699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	105469	57.519	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.040%
35) Dibromofluoromethane	5.292	113	79738	50.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.620%
50) Toluene-d8	7.899	98	309825	52.565	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.120%
62) 4-Bromofluorobenzene	10.636	95	117645	49.775	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.560%
Target Compounds						
16) Acetone	2.626	43	7367	7.893	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045588.D
Acq On : 06 Nov 2021 03:17
Operator : SY/MD
Sample : M4478-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TWP02

Quant Time: Nov 06 21:52:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

