

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045589.D
 Acq On : 06 Nov 2021 03:40
 Operator : SY/MD
 Sample : M4504-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB01-20211102

Quant Time: Nov 06 21:52:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	123986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	242874	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	244422	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	122204	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	106796	57.515	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.040%
35) Dibromofluoromethane	5.295	113	81555	50.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.580%
50) Toluene-d8	7.899	98	317277	52.619	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.240%
62) 4-Bromofluorobenzene	10.635	95	119732	49.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045589.D
Acq On : 06 Nov 2021 03:40
Operator : SY/MD
Sample : M4504-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB01-20211102

Quant Time: Nov 06 21:52:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

