

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045594.D
 Acq On : 06 Nov 2021 05:34
 Operator : SY/MD
 Sample : M4472-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-2-KD

Quant Time: Nov 06 21:54:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	120850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	235322	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	241772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	118032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	102605	56.692	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.380%
35) Dibromofluoromethane	5.292	113	78986	50.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.540%
50) Toluene-d8	7.899	98	313192	53.609	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.220%
62) 4-Bromofluorobenzene	10.632	95	115621	49.354	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.700%
Target Compounds						
						Qvalue
16) Acetone	2.626	43	15185	16.482	ug/l	99
20) Methylene Chloride	3.051	84	6286	3.927	ug/l	98
43) Isopropyl Acetate	5.912	43	15522	3.850	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045594.D
Acq On : 06 Nov 2021 05:34
Operator : SY/MD
Sample : M4472-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-2-KD

Quant Time: Nov 06 21:54:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

