

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
 Data File : VU045595.D
 Acq On : 06 Nov 2021 05:57
 Operator : SY/MD
 Sample : M4470-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SP-3

Quant Time: Nov 06 21:54:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	119170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	231974	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	241266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	118895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	103166	57.806	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 115.620%	
35) Dibromofluoromethane	5.295	113	78421	50.634	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.260%	
50) Toluene-d8	7.899	98	311205	54.037	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.080%	
62) 4-Bromofluorobenzene	10.635	95	116447	50.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.840%	
Target Compounds						Qvalue
16) Acetone	2.626	43	14750	16.235	ug/l	99
20) Methylene Chloride	3.054	84	6084	3.854	ug/l	94
43) Isopropyl Acetate	5.912	43	16335	4.110	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110521\
Data File : VU045595.D
Acq On : 06 Nov 2021 05:57
Operator : SY/MD
Sample : M4470-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SP-3

Quant Time: Nov 06 21:54:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

