

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027975.D
 Acq On : 06 Nov 2018 10:35
 Operator : MD/SY
 Sample : J5800-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 06 11:01:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	207368	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	388507	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	351155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	150636	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	169341	54.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.86%	
35) Dibromofluoromethane	4.89	113	127438	52.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.04%	
50) Toluene-d8	7.57	98	498912	52.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.44%	
62) 4-Bromofluorobenzene	10.31	95	170555	48.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027975.D
Acq On : 06 Nov 2018 10:35
Operator : MD/SY
Sample : J5800-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Nov 06 11:01:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

