

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027978.D
 Acq On : 06 Nov 2018 11:44
 Operator : MD/SY
 Sample : J5822-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-01

Quant Time: Nov 06 12:17:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	376230	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	341406	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	145133	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	164966	54.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.66%	
35) Dibromofluoromethane	4.89	113	122495	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	7.57	98	487317	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
62) 4-Bromofluorobenzene	10.31	95	164225	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027978.D
Acq On : 06 Nov 2018 11:44
Operator : MD/SY
Sample : J5822-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-01

Quant Time: Nov 06 12:17:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

