

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027981.D
 Acq On : 06 Nov 2018 12:54
 Operator : MD/SY
 Sample : J5822-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-6

Quant Time: Nov 06 17:23:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	374265	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	335349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	141786	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	164059	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	4.89	113	120340	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
50) Toluene-d8	7.57	98	478516	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	10.31	95	161888	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
Target Compounds						
11) Tert butyl alcohol	2.83	59	190635	238.955	ug/l	Qvalue 100
16) Acetone	2.32	43	15998	8.556	ug/l	94
19) Methyl tert-butyl Ether	3.00	73	20053	2.341	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027981.D
Acq On : 06 Nov 2018 12:54
Operator : MD/SY
Sample : J5822-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-6

Quant Time: Nov 06 17:23:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

