

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027982.D
 Acq On : 06 Nov 2018 13:17
 Operator : MD/SY
 Sample : J5822-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-7

Quant Time: Nov 06 13:38:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	370738	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	334948	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	145983	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	161809	54.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.32%	
35) Dibromofluoromethane	4.89	113	120368	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
50) Toluene-d8	7.57	98	477424	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
62) 4-Bromofluorobenzene	10.31	95	163350	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
Target Compounds						
16) Acetone	2.32	43	13295	7.239	ug/l	98
64) Tetrachloroethene	8.23	164	4652	2.142	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027982.D
Acq On : 06 Nov 2018 13:17
Operator : MD/SY
Sample : J5822-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-7

Quant Time: Nov 06 13:38:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

