

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027983.D
 Acq On : 06 Nov 2018 13:41
 Operator : MD/SY
 Sample : J5822-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-8

Quant Time: Nov 06 14:01:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	204650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	375312	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	338763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	141640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	165928	54.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.06%	
35) Dibromofluoromethane	4.89	113	124669	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
50) Toluene-d8	7.57	98	483837	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
62) 4-Bromofluorobenzene	10.31	95	163429	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
Target Compounds						
16) Acetone	2.33	43	2648	1.403	ug/l	91
64) Tetrachloroethene	8.23	164	3115	1.418	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027983.D
Acq On : 06 Nov 2018 13:41
Operator : MD/SY
Sample : J5822-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-8

Quant Time: Nov 06 14:01:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

