

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027971.D
 Acq On : 06 Nov 2018 08:51
 Operator : MD/SY
 Sample : VU1106WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1106WBL01

Quant Time: Nov 06 10:21:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	209233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	379446	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	332300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	130463	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	160000	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
35) Dibromofluoromethane	4.89	113	124651	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	
50) Toluene-d8	7.57	98	476636	51.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.14%	
62) 4-Bromofluorobenzene	10.31	95	156931	46.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.08%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027971.D
Acq On : 06 Nov 2018 08:51
Operator : MD/SY
Sample : VU1106WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1106WBL01

Quant Time: Nov 06 10:21:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. 0.00 min

Lab File: VU027971.D

Acq: 06 Nov 2018 08:51

Instrument :

MSVOA_U

ClientSampleId :


VU1106WBL01

Tot Ion:168 Resp: 209233

Ion Ratio Lower Upper

168 100

99 58.0 47.9 71.9



Abundance Ion 167.90 (167.60 to 168.60): VU027906.D (-1168) (-)

Ion 98.90 (98.60 to 99.60): VU027971.D (-1092) (-)

4.99

100000

80000



60000

40000

20000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 51.434 ug/l

RT: 5.31 min Scan# 1287

Delta R.T. -0.00 min

Lab File: VU027971.D


Acq: 06 Nov 2018 08:51

Tgt Ion: 65 Resp: 160000

Ion Ratio Lower Upper

65 100

67 54.8 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VU027906.D (-1270) (-)

Ion 66.90 (66.60 to 67.60): VU027971.D (-1194) (-)

5.31

80000


60000


40000

20000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.89 min Scan# 1466
 Delta R.T. -0.00 min
 Lab File: VU027971.D
 Acq: 06 Nov 2018 08:51

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1106WBL01

Tot Ion:114 Resp: 379446
 Ion Ratio Lower Upper
 114 100
 63 23.7 0.0 50.8
 88 16.5 0.0 35.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VU0
 Ion 87.90 (87.60 to 88.60): VU0






#35
 Dibromofluoromethane
 Concen: 52.101 ug/l
 RT: 4.89 min Scan# 1155
 Delta R.T. 0.00 min
 Lab File: VU027971.D
 Acq: 06 Nov 2018 08:51

Tgt Ion:113 Resp: 124651
 Ion Ratio Lower Upper
 113 100
 111 101.2 83.0 124.6
 192 19.7 15.6 23.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 51.573 ug/l
RT: 7.57 min Scan# 1988
Delta R.T. -0.00 min
Lab File: VU027971.D
Acq: 06 Nov 2018 08:51

Instrument :
MSVOA_U
ClientSampleId :
VU1106WBL01

Tot Ion: 98 Resp: 476636
Ion Ratio Lower Upper
98 100
100 63.4 50.1 75.1




Abundance Ion 98.00 (97.70 to 98.70): VU0
Ion 100.00 (99.70 to 100.70): VU






#62
4-Bromofluorobenzene
Concen: 46.037 ug/l
RT: 10.31 min Scan# 2841
Delta R.T. -0.00 min
Lab File: VU027971.D
Acq: 06 Nov 2018 08:51

Tgt Ion: 95 Resp: 156931
Ion Ratio Lower Upper
95 100
174 78.9 0.0 156.4
176 72.7 0.0 150.8



Abundance Ion 94.90 (94.60 to 95.60): VU0
Ion 173.80 (173.50 to 174.50): V
Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.09 min Scan# 2462
Delta R.T. -0.00 min
Lab File: VU027971.D
Acq: 06 Nov 2018 08:51

Instrument :
MSVOA_U
ClientSampleId :
VU1106WBL01

Tot Ion:117 Resp: 332300
Ion Ratio Lower Upper
117 100
82 60.4 48.6 73.0
119 32.5 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VU0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.48 min Scan# 3207
Delta R.T. -0.00 min
Lab File: VU027971.D
Acq: 06 Nov 2018 08:51

Tgt Ion:152 Resp: 130463
Ion Ratio Lower Upper
152 100
115 60.1 43.0 129.0
150 158.0 0.0 349.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

