

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110618\
 Data File : VU027977.D
 Acq On : 06 Nov 2018 11:21
 Operator : MD/SY
 Sample : J5822-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 06 11:57:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	205288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	379659	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	338314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	142745	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	163701	53.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.28%	
35) Dibromofluoromethane	4.89	113	125006	52.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.44%	
50) Toluene-d8	7.57	98	484606	52.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.82%	
62) 4-Bromofluorobenzene	10.31	95	161144	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110618\
Data File : VU027977.D
Acq On : 06 Nov 2018 11:21
Operator : MD/SY
Sample : J5822-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 06 11:57:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. 0.00 min

Lab File: VU027977.D

Acq: 06 Nov 2018 11:21

Instrument :

MSVOA_U

ClientSampleId :


TRIP-BLANK

Tot Ion:168 Resp: 205288




Ion Ratio Lower Upper

168 100

99 59.0 47.9 71.9



Abundance Ion 167.90 (167.60 to 168.60): VU027906.D (-1168) (-)



#33

1,2-Dichloroethane-d4

Concen: 53.635 ug/l

RT: 5.31 min Scan# 1287

Delta R.T. -0.00 min

Lab File: VU027977.D


Acq: 06 Nov 2018 11:21

Tgt Ion: 65 Resp: 163701



Ion Ratio Lower Upper


65 100

67 55.1 0.0 109.0



Abundance Ion 64.90 (64.60 to 65.60): VU027906.D (-1270) (-)









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 5.89 min Scan# 1467
Delta R.T. 0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Tot Ion:114 Resp: 379659
Ion Ratio Lower Upper
114 100
63 24.8 0.0 50.8
88 17.2 0.0 35.8




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VU0
Ion 87.90 (87.60 to 88.60): VU0






#35
Dibromofluoromethane
Concen: 52.221 ug/l
RT: 4.89 min Scan# 1155
Delta R.T. 0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21

Tgt Ion:113 Resp: 125006
Ion Ratio Lower Upper
113 100
111 101.9 83.0 124.6
192 18.9 15.6 23.4



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.406 ug/l
RT: 7.57 min Scan# 1988
Delta R.T. -0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21




Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK


Tot Ion: 98 Resp: 484606
Ion Ratio Lower Upper
98 100
100 63.2 50.1 75.1



#62
4-Bromofluorobenzene
Concen: 47.246 ug/l
RT: 10.31 min Scan# 2841
Delta R.T. -0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21

Tgt Ion: 95 Resp: 161144
Ion Ratio Lower Upper
95 100
174 78.7 0.0 156.4
176 75.4 0.0 150.8






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.09 min Scan# 2462
Delta R.T. -0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Tot Ion	Ratio	Lower	Upper
117	100		
82	60.5	48.6	73.0
119	31.7	25.8	38.8






Abundance

Ion 116.90 (116.60 to 117.60): V


Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.48 min Scan# 3207
Delta R.T. -0.00 min
Lab File: VU027977.D
Acq: 06 Nov 2018 11:21

Tgt Ion	Ratio	Lower	Upper
152	100		
115	59.5	43.0	129.0
150	157.0	0.0	349.4



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

