

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU110619\
 Data File : VU035653.D
 Acq On : 06 Nov 2019 17:26
 Operator : JC/SP
 Sample : K5716-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BE887

Quant Time: Nov 07 06:27:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 07 04:46:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	627595	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	606802	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	177898	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	169734	33.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.98%
7) Chloroethane-d5	1.93	69	158400	39.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.26%
11) 1,1-Dichloroethene-d2	2.60	63	229342	29.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.18%#
21) 2-Butanone-d5	4.69	46	286812	96.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.87%
24) Chloroform-d	5.11	84	352329	45.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.02%
26) 1,2-Dichloroethane-d4	5.75	65	221022	46.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.18%
32) Benzene-d6	5.77	84	724118	45.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.56%
36) 1,2-Dichloropropane-d6	6.74	67	242329	47.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.48%
41) Toluene-d8	7.93	98	646741	43.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.70%
43) trans-1,3-Dichloropropene-	8.22	79	105509	44.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.30%
47) 2-Hexanone-d5	8.68	63	177908	90.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.50%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	338863	46.45	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.90%
64) 1,2-Dichlorobenzene-d4	12.22	152	180909	50.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.84%

Target Compounds

					Qvalue
13) Acetone	2.67	43	9982	3.832 ug/L	97
34) Trichloroethene	6.58	95	78645	18.140 ug/L	98
46) Tetrachloroethene	8.58	164	30444	8.398 ug/L	98
53) m,p-Xylene	9.72	106	13714	1.913 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU110619\
Data File : VU035653.D
Acq On : 06 Nov 2019 17:26
Operator : JC/SP
Sample : K5716-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BE887

Quant Time: Nov 07 06:27:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 07 04:46:39 2019
Response via : Initial Calibration

