

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U110620\
 Data File : VU041121.D
 Acq On : 06 Nov 2020 18:35
 Operator : SY/MD
 Sample : L4648-01DL 10X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AA8DL

Quant Time: Nov 06 23:56:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 06 23:04:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	66472	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	68950	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	30808	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	21455	3.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.80%
7) Chloroethane-d5	1.92	69	19372	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	2.58	63	35271	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	4.66	46	100562	54.68	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.36%
24) Chloroform-d	5.08	84	50409	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
26) 1,2-Dichloroethane-d4	5.72	65	31926	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	5.74	84	92127	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.70	67	31932	5.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.40%
41) Toluene-d8	7.91	98	78382	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
43) trans-1,3-Dichloropropene-	8.19	79	12740	4.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.80%
46) 2-Hexanone-d5	8.64	63	74722	53.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.66%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	28760	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	32153	5.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.00%

Target Compounds					Ovalue
9) Trichlorofluoromethane	2.15	101	18166	2.141 ug/L	99
16) Methylene chloride	3.06	84	1187	0.217 ug/L	97
47) Tetrachloroethene	8.56	164	30224	7.840 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110620\
Data File : VU041121.D
Acq On : 06 Nov 2020 18:35
Operator : SY/MD
Sample : L4648-01DL 10X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AA8DL

Quant Time: Nov 06 23:56:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 06 23:04:51 2020
Response via : Initial Calibration

