

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110620\
 Data File : VU041128.D
 Acq On : 06 Nov 2020 21:18
 Operator : SY/MD
 Sample : L4646-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB0G1

Quant Time: Nov 07 00:23:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 06 23:04:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	57318	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	60057	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25615	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	18772	3.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.00%
7) Chloroethane-d5	1.92	69	17338	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.58	63	30723	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	4.65	46	82942	52.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.60%
24) Chloroform-d	5.08	84	44535	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.72	65	27832	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.74	84	78893	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.70	67	28485	5.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.00%
41) Toluene-d8	7.91	98	66134	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	8.19	79	10441	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.64	63	57874	47.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.84%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	23896	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	27193	6.05	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	121.00%#

Target Compounds

					Ovalue
13) Acetone	2.66	43	3348	3.088 ug/L	72
22) cis-1,2-Dichloroethene	4.69	96	3693	0.870 ug/L	91
34) Trichloroethene	6.55	95	3503	0.767 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110620\
Data File : VU041128.D
Acq On : 06 Nov 2020 21:18
Operator : SY/MD
Sample : L4646-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB0G1

Quant Time: Nov 07 00:23:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 06 23:04:51 2020
Response via : Initial Calibration

