

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110623\
 Data File : VU056084.D
 Acq On : 06 Nov 2023 14:50
 Operator : MD/SY
 Sample : 05239-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP8

Quant Time: Nov 07 04:09:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 07 04:07:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	352493	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	344624	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	153596	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	84865	3.043	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.800%
7) Chloroethane-d5	1.914	69	74453	3.425	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.600%
11) 1,1-Dichloroethene-d2	2.566	65	40730	3.441	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.800%
20) 2-Butanone-d5	4.631	46	231231	46.484	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.960%
24) Chloroform-d	5.062	84	201971	3.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
26) 1,2-Dichloroethane-d4	5.698	65	95611	3.798	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.000%
32) Benzene-d6	5.724	84	390918	3.808	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.200%
36) 1,2-Dichloropropane-d6	6.685	67	124792	4.101	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.000%
41) Toluene-d8	7.894	98	324918	3.489	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.800%#
43) trans-1,3-Dichloroprop...	8.177	79	44627	3.749	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.631	63	171286	56.692	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.380%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	92600	4.178	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	116853	4.078	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	81.600%
Target Compounds						
					Qvalue	
13) Acetone	2.647	43	35766	12.617	ug/L	95
21) 2-Butanone	4.730	43	11717	2.812	ug/L	89
25) Chloroform	5.084	83	86661	2.350	ug/L	98
38) Bromodichloromethane	7.106	83	13426	0.530	ug/L	91
49) Dibromochloromethane	8.807	129	1908	0.122	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110623\
Data File : VU056084.D
Acq On : 06 Nov 2023 14:50
Operator : MD/SY
Sample : 05239-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP8

Quant Time: Nov 07 04:09:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 07 04:07:55 2023
Response via : Initial Calibration

