

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU027992.D
 Acq On : 07 Nov 2018 09:32
 Operator : MD/SY
 Sample : VU1107WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1107WBL01

Quant Time: Nov 07 10:04:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	205254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	391000	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	352557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	140128	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	171337	56.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.30%	
35) Dibromofluoromethane	4.88	113	129894	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
50) Toluene-d8	7.57	98	506228	53.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.32%	
62) 4-Bromofluorobenzene	10.31	95	166996	47.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU027992.D
Acq On : 07 Nov 2018 09:32
Operator : MD/SY
Sample : VU1107WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1107WBL01

Quant Time: Nov 07 10:04:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

