

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU028000.D
 Acq On : 07 Nov 2018 13:03
 Operator : MD/SY
 Sample : PB114573ZHE#05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114573ZHE#05

Quant Time: Nov 07 14:23:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	379819	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	335989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	139767	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	167410	56.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.94%	
35) Dibromofluoromethane	4.89	113	123626	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
50) Toluene-d8	7.57	98	487860	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
62) 4-Bromofluorobenzene	10.31	95	163704	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
Target Compounds						
16) Acetone	2.32	43	10921	5.938	ug/l	96
18) Methyl Acetate	2.62	43	5965	1.179	ug/l	93
20) Methylene Chloride	2.70	84	12609	4.695	ug/l	95
43) Isopropyl Acetate	5.55	43	240751	26.364	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU028000.D
Acq On : 07 Nov 2018 13:03
Operator : MD/SY
Sample : PB114573ZHE#05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114573ZHE#05

Quant Time: Nov 07 14:23:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

