

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU028002.D
 Acq On : 07 Nov 2018 13:49
 Operator : MD/SY
 Sample : J5839-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-7-D

Quant Time: Nov 07 14:27:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	199098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	379885	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	343377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	149026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	167898	56.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.44%	
35) Dibromofluoromethane	4.88	113	123241	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
50) Toluene-d8	7.57	98	494882	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
62) 4-Bromofluorobenzene	10.31	95	167582	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
Target Compounds						
16) Acetone	2.32	43	10640	5.793	ug/l	95
18) Methyl Acetate	2.62	43	5313	1.052	ug/l	97
20) Methylene Chloride	2.70	84	11378	4.243	ug/l	96
43) Isopropyl Acetate	5.55	43	227477	24.906	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU028002.D
Acq On : 07 Nov 2018 13:49
Operator : MD/SY
Sample : J5839-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-7-D

Quant Time: Nov 07 14:27:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

