

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU028003.D
 Acq On : 07 Nov 2018 14:12
 Operator : MD/SY
 Sample : J5839-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-7-S

Quant Time: Nov 07 14:50:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	195573	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	371392	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	330640	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	136213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	161533	55.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.10%	
35) Dibromofluoromethane	4.89	113	120639	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
50) Toluene-d8	7.57	98	478253	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
62) 4-Bromofluorobenzene	10.31	95	157810	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
Target Compounds						
16) Acetone	2.32	43	12199	6.762	ug/l	94
18) Methyl Acetate	2.62	43	6537	1.317	ug/l #	91
20) Methylene Chloride	2.70	84	12193	4.629	ug/l	97
43) Isopropyl Acetate	5.55	43	225307	25.233	ug/l	99
95) Naphthalene	13.75	128	27042	2.247	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU028003.D
Acq On : 07 Nov 2018 14:12
Operator : MD/SY
Sample : J5839-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-7-S

Quant Time: Nov 07 14:50:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

