

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU028006.D
 Acq On : 07 Nov 2018 15:22
 Operator : MD/SY
 Sample : J5762-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BU-02-110518

Quant Time: Nov 07 16:12:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190748	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	367957	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	327165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	137891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	159926	56.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.78%	
35) Dibromofluoromethane	4.89	113	118917	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
50) Toluene-d8	7.57	98	472740	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
62) 4-Bromofluorobenzene	10.31	95	158775	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
Target Compounds						
16) Acetone	2.32	43	13630	7.746	ug/l	99
18) Methyl Acetate	2.62	43	5136	1.061	ug/l	100
20) Methylene Chloride	2.70	84	17237	6.710	ug/l	95
43) Isopropyl Acetate	5.55	43	225361	25.474	ug/l	99
68) m/p-Xylenes	9.38	106	5326	1.154	ug/l	99
95) Naphthalene	13.75	128	77837	6.725	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU028006.D
Acq On : 07 Nov 2018 15:22
Operator : MD/SY
Sample : J5762-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BU-02-110518

Quant Time: Nov 07 16:12:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

