

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110718\
 Data File : VU028007.D
 Acq On : 07 Nov 2018 15:45
 Operator : MD/SY
 Sample : J5762-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BU-01-110518

Quant Time: Nov 07 16:14:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	193685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	369780	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	334171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	140216	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	164654	57.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.36%	
35) Dibromofluoromethane	4.89	113	120917	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
50) Toluene-d8	7.57	98	484369	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
62) 4-Bromofluorobenzene	10.31	95	157444	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
Target Compounds						
16) Acetone	2.32	43	11656	6.524	ug/l	97
18) Methyl Acetate	2.62	43	5446	1.108	ug/l	94
20) Methylene Chloride	2.70	84	22684	8.696	ug/l	97
43) Isopropyl Acetate	5.55	43	210902	23.723	ug/l	100
95) Naphthalene	13.75	128	14064	1.045	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110718\
Data File : VU028007.D
Acq On : 07 Nov 2018 15:45
Operator : MD/SY
Sample : J5762-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BU-01-110518

Quant Time: Nov 07 16:14:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

