

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110719\
 Data File : VU035661.D
 Acq On : 07 Nov 2019 15:56
 Operator : JC/SP
 Sample : VU1107WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK50

Quant Time: Nov 08 06:26:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 08 06:13:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	708419	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	670992	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	201969	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	264925	46.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.60%
7) Chloroethane-d5	1.93	69	223471	48.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.82%
11) 1,1-Dichloroethene-d2	2.59	63	314408	35.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.66%
21) 2-Butanone-d5	4.69	46	351220	105.09	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.09%
24) Chloroform-d	5.11	84	420661	47.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.75	65	259695	47.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.96%
32) Benzene-d6	5.77	84	894535	51.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.28%
36) 1,2-Dichloropropane-d6	6.73	67	289020	50.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.92%
41) Toluene-d8	7.93	98	812847	49.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.54%
43) trans-1,3-Dichloropropene-	8.21	79	126515	48.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.84%
47) 2-Hexanone-d5	8.68	63	227258	104.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.54%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	386832	47.96	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.92%
64) 1,2-Dichlorobenzene-d4	12.22	152	208524	51.19	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.38%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110719\
Data File : VU035661.D
Acq On : 07 Nov 2019 15:56
Operator : JC/SP
Sample : VU1107WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK50

Quant Time: Nov 08 06:26:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 08 06:13:27 2019
Response via : Initial Calibration

