

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U110719\
 Data File : VU035670.D
 Acq On : 07 Nov 2019 19:22
 Operator : JC/SP
 Sample : K5739-03DL 5X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CR1DL

Quant Time: Nov 08 07:22:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 08 06:13:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	703018	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	684097	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	204007	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.62	65	259465	45.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.40%
7) Chloroethane-d5	1.93	69	219817	48.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.96%
11) 1,1-Dichloroethene-d2	2.60	63	309795	35.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.16%
21) 2-Butanone-d5	4.69	46	318899	96.15	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.15%
24) Chloroform-d	5.11	84	416170	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.92%
26) 1,2-Dichloroethane-d4	5.75	65	258665	48.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.30%
32) Benzene-d6	5.77	84	882231	49.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.94%
36) 1,2-Dichloropropane-d6	6.74	67	287058	49.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.28%
41) Toluene-d8	7.93	98	808885	48.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.18%
43) trans-1,3-Dichloropropene-	8.21	79	124425	46.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.42%
47) 2-Hexanone-d5	8.68	63	197136	88.95	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.95%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	376859	45.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.66%
64) 1,2-Dichlorobenzene-d4	12.22	152	211972	51.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.04%

Target Compounds

					Ovalue
13) Acetone	2.67	43	9715	3.329 ug/L	97
63) 1,4-Dichlorobenzene	11.87	146	8011	1.225 ug/L	89
65) 1,2-Dichlorobenzene	12.24	146	43460	6.633 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110719\
Data File : VU035670.D
Acq On : 07 Nov 2019 19:22
Operator : JC/SP
Sample : K5739-03DL 5X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CR1DL

Quant Time: Nov 08 07:22:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM103119WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 08 06:13:27 2019
Response via : Initial Calibration

