

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
 Data File : VU056095.D
 Acq On : 07 Nov 2023 13:10
 Operator : MD/SY
 Sample : VU1107WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK060

Quant Time: Nov 08 00:38:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 00:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	387272	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	388568	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	167313	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	143295	5.961	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 119.200%	
7) Chloroethane-d5	1.907	69	109951	6.740	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 134.800%#	
11) 1,1-Dichloroethene-d2	2.563	65	61157	5.775	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 115.400%	
20) 2-Butanone-d5	4.618	46	258106	65.450	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 130.900%#	
24) Chloroform-d	5.055	84	263084	5.749	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 115.000%	
26) 1,2-Dichloroethane-d4	5.695	65	123744	6.142	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 122.800%	
32) Benzene-d6	5.721	84	553646	5.933	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 118.600%	
36) 1,2-Dichloropropane-d6	6.685	67	165588	6.028	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 120.600%	
41) Toluene-d8	7.894	98	481503	5.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 112.600%	
43) trans-1,3-Dichloroprop...	8.177	79	60729	5.783	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 115.600%	
46) 2-Hexanone-d5	8.631	63	209104	60.755	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 121.500%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	117689	6.002	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 120.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	165585	6.418	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 128.400%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
Data File : VU056095.D
Acq On : 07 Nov 2023 13:10
Operator : MD/SY
Sample : VU1107WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK060

Quant Time: Nov 08 00:38:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 00:38:11 2023
Response via : Initial Calibration

